



Instrument
Markets

UV Measurement and Process Control Instruments

EIT LLC 309 Kelly's Ford Plaza SE Leesburg, VA 20175 USA

Phone: (703) 478-0700 Email: uv@eit.com Web: www.eit.com

Subject: Counterfeiting of EIT Instruments

Date: November 2020

Counterfeiting of EIT's UviCure Plus® II and Power Puck® II instruments has been confirmed in China. After an investigation, EIT is pursuing all legal remedies to stop the counterfeiting and use of EIT's registered trademarks.

Counterfeit units or units suspected of being counterfeit should be immediately removed from service. Electronic Instrumentation and Technology LLC (EIT LLC) cannot guarantee the performance of counterfeit units and is not responsible for results and products manufactured with them.

EIT instruments are critical to the manufacturing, quality assurance and safety of UV cured products in all types of applications. From a functional, reliability and performance standpoint, the counterfeit units are inferior to a genuine EIT unit.

Examination of the counterfeit units showed extreme quality differences:

- **Physically:** The counterfeit housing 'looks' similar but the anodized coating used will not withstand repeated UV exposures
- **Optically:** The optics in the units examined are old, recycled and/or damaged
- **Electronically:** The quality of the components, printed circuit board and workmanship including the solder and solder joints are not up to accepted industry standards
- **Calibration:** The units were not calibrated using EIT's proprietary standards and calibration processes.

Contact EIT at uv@eit.com or 703-478-0700 if you suspect a counterfeit unit. We need a clear high-resolution image, unit type, serial number, date and place of purchase and to determine if the unit is genuine or counterfeit.